

Myocardial Imaging Tissue Doppler and Strain Imaging

Steven J. Lester MD, FRCP(C), FACC, FASE

@2017 MFMER | 368226

DISCLOSURE


Relevant Financial Relationship(s)


None


Off Label Usage


None




































Acoustic pattern tracking Speckle Tracking

Velocity is estimated as a shift of each object divided by time between successive frames (or multiplied by Frame Rate)-->

2D vector: (Vx, Vy) = (dX, dY) * FR



MAYO CLINIC


©2017 MFMER | 3682262-2


Doppler Independent Techniques (Speckle Tracking)


Potential Advantage?


- Signal noise
- Speckle tracking by principle is angle independent
- Gray scale (standard views)
- Monitor strain in two rather than one dimension
- Minimal user input
- Assessment of rotation: derived from circumferential strain at different levels in the heart (NO fixed sample volume)


MAYO CLINIC














Echocardiographic Measures of Myocardial Deformation by Speckle-Tracking Technologies: The Need for Standardization?

Matthew R. Nelson, MD, R. Todd Hurst, MD, Serageldin F. Raslan, MD, Stephen Cha, MS, Susan Wilansky, MD, and Steven J. Lester, MD, Scottsdale, Arizona; Rochester, Minnesota

Echocardiographic Measures of Myocardial Deformation by Speckle-Tracking Technologies: The Need for Standardization?

than two left ventricular endocardial segments poorly delineated were excluded. GLS was obtained from the apical four-chamber, three-chamber, and two-chamber views using two independent speckle-tracking echocardiographic software packages (Echolnsight version 1.5.0 and Image-Arena version 4.5). Linear regression analysis and paired t tests were used to compare GLS results. Intraclass correlation coefficients and Bland-Altman plots were used for assessments of reliability.

Results: The "out-of-the-box" mean GLS was $-12.99 \pm 2.38\%$ using Echolnsight and $-16.87 \pm 2.84\%$ using Image-Arena (mean difference, $3.87 \pm 2.42\%$, P = .0001). Agreement between the software packages was moderate (intraclass correlation coefficient, 0.43;95% confidence interval, 0.32-0.55). Using uniform variables to derive 0.32 ± 0.32 . The peak segmental peak segmental peak segmental 0.32 ± 0.32 . The peak segmental 0.32 ± 0.32 and 0.32 ± 0.32 .

J Am Soc Echocardiogr 2012:25:1189-94

% (mean differval, 0.52–0.79).

Conclusions: Image-Arena GLS results were consistently different (more negative) than EchoInsight measures out of the box but became similar when information used to derive GLS was uniform. The evolution of measures of myocardial mechanics into routine clinical practice will require vigilance and standardization of the various techniques, necessitating independent validation of commercially available speckle-tracking echocardiographic products. [J Am Soc Echocardiogr 2012;25:1189-94.]

Keywords: Speckle-tracking, Strain, Echocardiography

©2017 MFMER | 3682262-



REPRODUCIBILITY OF LEFT VENTRICULAR STRAIN

Head-to-Head Comparison of Global Longitudinal Strain Measurements among Nine Different Vendors The EACVI/ASE Inter-Vendor Comparison Study

Konstantinos E. Farsalinos, MD, Ana M. Daraban, MD, Serkan Ünlü, MD, James D. Thomas, MD, Luigi P. Badano, MD, PhD, and Jens-Uwe Voigt, MD, PhD, Leuven, Belgium; Chicago, Illinois; and Padua, Italy


Background: This study was planned by the EACVI/ASE/Industry Task Force to Standardize Deformation Imaging to (1) test the variability of speckle-tracking global longitudinal strain (GLS) measurements among different vendors and (2) compare GLS measurement variability with conventional echocardiographic parameters.

Methods: Sixty-two volunteers were studied using ultrasound systems from seven manufacturers. Each volunteer was examined by the same sonographer on all machines. Inter- and intraobserver variability was determined in a true test-retest setting. Conventional echocardiographic parameters were acquired for comparison. Using the software packages of the respective manufacturer and of two software-only vendors, endocardial GLS was measured because it was the only GLS parameter that could be provided by all manufactures. We compared GLS_{AV} (the average from the three apical views) and GLS_{4CH} (measured in the four-chamber view) measurements among vendors and with the conventional echocardiographic parameters.

Results: Absolute values of GLS_{AV} ranged from 18.0% to 21.5%, while GLS_{4CH} ranged from 17.9% to 21.4%. The absolute difference between vendors for GLS_{AV} was up to 3.7% strain units (P < .001). The interobserver relative mean errors were 5.4% to 8.6% for GLS_{AV} and 6.2% to 11.0% for GLS_{4CH} , while the intraobserver relative mean errors were 4.9% to 7.3% and 7.2% to 11.3%, respectively. These errors were lower than for left ventricular ejection fraction and most other conventional echocardiographic parameters.

Conclusion: Reproducibility of GLS measurements was good and in many cases superior to conventional echocardiographic measurements. The small but statistically significant variation among vendors should be considered in performing serial studies and reflects a reference point for ongoing standardization efforts. (J Am Soc Echocardiogr 2015;28:1171-81.)





GUIDELINES AND STANDARDS

Recommendations for Cardiac Chamber

Global Longitudinal Peak Systolic Strain (GLS) "in the range of -20%"


are: Roberto M. Lang. MD. FASE, et al.


- "Optimize image quality, maximize frame rate and minimize foreshortening".
- "When regional tracking is suboptimal in more than two myocardial segments in a single view the calculation of GLS should be avoided".


assembled by the American Society of Echocardiography and the European Association of Cardiovascular Imaging. This document provides updated normal values for all four cardiac chambers, including three-dimensional echocardiography and myocardial deformation, when possible, on the basis of considerably larger numbers of normal subjects, compiled from multiple databases. In addition, this document attempts to eliminate several minor discrepancies that existed between previously published guidelines. (J Am Soc Echocardiogr 2015;28:1-39.)


Keywords: Adult echocardiography, Transthoracic echocardiography, Ventricular function, Normal values






































Case

- 59-year-old male
- Acute Myeloid Leukemia
- No prior history of vascular disease.
- Hypertension treated with Amlodipine.
- About to begin chemotherapy based treatment

MAYO CLINIC










Niccolo Machiavelli (1469-1527)

"...at the beginning a disease is easy to cure but difficult to diagnose; but as time passes, not having been recognized or treated at the outset, it becomes easy to diagnose but difficult to cure."



MAYO CLINIC





Case

- 59-year-old male
- Acute Myeloid Leukemia
- No prior history of vascular disease.
- Hypertension treated with Amlodipine.
- About to begin chemotherapy based treatment



©2017 MFMER | 3682262-5

Baseline Echocardiogram

LVEF = 66%, EDVI = 53 ml/m^2








Baseline Echocardiogram

Global Longitudinal Peak Systolic Strain

LVEF = 0616P% SS Avg = -17.3%



MAYO CLINIC

2017 MFMER | 3682262-5


EXPERT CONSENSUS STATEMENT


Expert Consensus for Multimodality Imaging
Evaluation of Adult Patients during and after Cancer
Therapy: A Report from the American Society of
Echocardiography and the European Association of
Cardiovascular Imaging


- 1. CTRCD if decrease in LVEF >10% to a value <53%
- 2. In patients with available baseline strain measurements, a relative percentage reduction of GLS of <8% from baseline appears not to be meaningful, and those >15% from baseline are very likely to be abnormal.

(J Am Soc Echocardiogr 2014;27:911-39.)

MAYO CLINIC










Case


- 64 year old woman
- HER2 positive infiltrating lobular carcinoma of the right breast
- HER2 positive ductal carcinoma insitu of the left breast.
- Preoperative chemotherapy with paclitaxel (80mg/m²) and trastuzumab. Paclitaxel discontinued after 8 infusions due to toxicity (neuropathy).
- Then preoperatively started Q3weekly doxorubicin/cyclophosphamide (discontinued after 2 cycles due fatigue and anorexia).

MAYO CLINIC









What should we do now?



- LVEF dropped from 65% to 59% (9% RRR)
- GLPSS dropped from -24% to -17% (29% RRR)
- Started treatment with Coreg and Enalapril
- Initiated adjuvant trastuzumab and anastrozole
- Serial echocardiograms Q2-3 months

©2017 MFMER | 3682262-

MAYO CLINIC


Completion of 1 year of adjuvant trastuzumab


LVEF = 59%






MAYO CLINIC











LV Mechanics in Mitral and Aortic Valve Diseases



Value of Functional Assessment Beyond Ejection Fraction


Elena Galli, MD, PhD,* Patrizio Lancellotti, MD, PhD,† Partho P. Sengupta, MD, DM,‡ Erwan Donal, MD, PhD*


ABSTRACT


- "LV dysfunction is frequently subclinical despite a normal ejection fraction. It may preceded the onset of symptoms and portend a poor outcome..."
- "The advent of novel tissue-tracking echo techniques has unleashed new opportunities for the clinical identification of early abnormalities in LV function".

regarding the use of these techniques to assess myocardial deformation in patients with valvular heart disease. (J Am Coll Cardiol Img 2014;7:1151-66) © 2014 by the American College of Cardiology Foundation.

MAYO CLINIC









Echocardiographic Evaluation of Aortic Stenosis


Rule #7:


The evaluation of left ventricular function should include not only a measure of ejection fraction but also global longitudinal strain.

MAYO CLINIC









Myocardial Imaging Proven Utility & Potential

A Masterpiece in Echocardiography?



- Subclinical LV dysfunction
- 2. HCM Phenocopies
- 3. Valve Disease
- 4. ...
- 5. ...

MAYO CLINIC

©2017 MFMER | 3682262-128



